


PULSE GENERATOR

PERFORMANCE CHECK

Model: ANH-HV1-N-C-1M3A


S.N.: 5567

Date: SEPT 11 1990



- a) Output signal amplitude:
No 1: $\approx -100V$
No 2: 0 TO $-150V$
- b) Pulse width:
2 NS (FWHM)
- c) Rise time:
 $\leq 1 NS$
- d) Fall time:
 $\leq 1 NS$
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

(A) 60 db ATTEN
-100 VOLTS/DIV
DELAY 1-2 $\approx 8 NS$



- h) PULSE 1-2 DELAY:
0 TO 8 NS.

(B) AS (A) BUT DELAY 1-2
 $\approx 0 NS$.