



PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVR-A-1-PW-P-PS-4NA  
S.N.: 5569  
Date: NOV 6 1990



Ⓐ  $R_L = 50\Omega$   
PRF = 20 KHz



Ⓑ (RISE TIME)  
100 VOLTS/DIV

- a) Output signal amplitude:  
0 TO +200 VOLTS
- b) Pulse width:  
0.1 TO 1.0  $\mu$ S
- c) Rise time:  
 $\leq 2$  NS
- d) Fall time:  
 $\leq 10$  NS
- e) PRF:  
0 TO 20 KHz\*
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 V  
50-60 KHz

\* CAUTION DO NOT  
EXCEED 20 KHz  
AS UNIT MAY  
FAIL!!

*[Signature]*