


PULSE GENERATOR

PERFORMANCE CHECK


Model: ANR-132-C-W-PN-EP-45CA3

S.N.: 5608

Date: SEPT 26 1990



P_{out} $R_L = 50 \Omega$
20 VOLT/DIV
PRF = 10 KHz



N_{out} $R_L = 50 \Omega$
20 VOLT/DIV
PRF = 10 KHz

- a) Output signal amplitude:
0 TO ± 50.0048
- b) Pulse width: TO 50ns
0.1 TO 5.0 μS
- c) Rise time:
 ≤ 5 NS
- d) Fall time:
 ≤ 5 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
5.0-6.0 Hz

[Handwritten signature]