


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVI-V-HUB-C-PN

S.N.: 5609

Date: SEPT 17 1940



a) Output signal amplitude:
0 TO ± 100 VOLTS

b) Pulse width:
4 TO 100 NS

c) Rise time:
 ≤ 200 PS


d) Fall time:
 ≤ 3 NS

e) PRF:
0 TO 5 KHZ

f) Jitter, stability:
OK

g) Prime power:
120/240 V
50-60 HZ

(A) 60 db ATTEN
100 VOLTS/DIV
PRF \approx 5 KHZ



(B) AS (A) BUT 100 NS/DIV
(RISE TIME)