


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-P

S.N.: 5616

Date: AUG 23, 1990



50dB ATTEN
∴ 32 VOLTS/DIV
PRF ≈ 50kHz

- a) Output signal amplitude:
0 TO +40VOLTS
- b) Pulse width:
2 TO 100ns
- c) Rise time:
≤ 350ps
- d) Fall time:
≤ 2ns
- e) PRF:
MIN PW: 0 TO 1MHz (see Fig. 1+2)
MAX PW: 0 TO 20kHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLT
150mA MAX

[Handwritten signature]