


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVA-14V1-P*

S.N.: *5636*

Date: *OCT 5 1990*



50 db ATTEN
32 VOLTS/DIV
PRF ≈ 70 KHz

- a) Output signal amplitude: *0 TO +100 VOLTS*
- b) Pulse width: *TO 50 ns*
≤ 1 NS
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 100 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 ± 1.0 VOLTS*
150 mA (max)