


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVN-1-P*
S.N.: *5646.*
Date: *OCT 23 1990*



40 dB ATTEN
5 V/DIV

- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.2 TO 1.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 100 PS
- e) PRF:
50 TO 250 MPD.
- f) Jitter, stability:
OK
- g) Prime power:
+24 VOLT.
500 mA (MAX)

[Signature]