


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-3A-P-IM*

S.N.: *5648*

Date: *NOV 7 1990*



- a) Output signal amplitude:  
*+25 TO +75 AMPS*
- b) Pulse width:  
*2.5 NSGZ*
- c) Rise time: *(1/2 Power)*  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 20 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*R<sub>L</sub> = 4.0 Ohms*

*60 db ATTEN*

*100 VOLTS/DIV*

*25 AMPS/DIV*

*+15 VOLTS*

*300 mA (MAX)*

*[Signature]*