


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-2W-C-M-PN*

S.N.: *5662*

Date: *NOV 13 1990*



~ 6 AMP S/DIS
PRF = 20 KHz

- a) Output signal amplitude:
0 TO ±10 AMP
- b) Pulse width: (*TO 5n*)
2 TO 50 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 1 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
150/240 V
50-60 Hz