


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMP-AN-2-ECL-P*

S.N.: *5673*

Date: *OCT 24 90*



*40 db ATTEN*  
*10 VOLT/DIV*  
*ABF = 50 KHZ*

- a) Output signal amplitude:  
*0 TO +10 VOLTS*
- b) Pulse width:  
*5 TO 50 NS*
- c) Rise time:  
*≤ 100 PS*
- d) Fall time:  
*≤ 100 PS*
- e) PRF:  
*0 TO 50 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 VOLT*  
*400 mA MA*

h) INPUT TECHNIQUE  
*ECL.*