


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVD-5-C-M-PN-E441*

S.N.: *5689*

Date: *NOV 28 1990*



- a) Output signal amplitude: *0 TO ± 350 VOLTS*
- b) Pulse width: *TO 50 ns*
(7 Amp)
- c) Rise time: *6 TO 100 NS*
≤ 3 NS
- d) Fall time: *≤ 4 NS*
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability: *OK*

*PHOTO OF ME
OUTPUT OF
AVD-51-E441-TO-5
(WITH 50 dB ATTEN)*

≈ 32 VOLTS

I_D = 0.2 x 32 AMPS

g) Prime power:
120/240 V
50-60 Hz