


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MR-B2-W-PN-PS-ND1*

S.N.: *5698*

Date: *NOV 30 1990*



- a) Output signal amplitude: *0 TO ± 50 VOLTS*
- b) Pulse width: *TO 50 μs*
0.1 TO 5.0 μs.
- c) Rise time: *≤ 5 NS*
- d) Fall time: *≤ 5 NS*
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 Hz

20 VOLTS/DIV
0.5 USEC/DIV
PRF = 10 KHz
R_L = 50 Ω.