


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-3-PN-PS-R03*


S.N.: *5703*

Date: *JAN 22 1991*



- a) Output signal amplitude:
0 TO ± 150 VOLTS
- b) Pulse width: *R_c > 1K.*
0.1^{MS} TO 1.0 MS
- c) Rise time:
≤ 30 NS
- d) Fall time:
≤ 30 NS
- e) PRF:
NA
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 HZ

ⓐ 5 MS/DIV PRF = 30 KHZ
50 VOLTS/DIV



ⓑ 1 US/DIV PRF = 2 KHZ
50 VOLTS/DIV

[Signature]