


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PCR-4B-PW-C-PN*


S.N.: *5705*

Date: *JAN 10 1991*



- a) Output signal amplitude:
0 TO ± 400 VOLTS
- b) Pulse width: *(0.5 to 5)*
0.1 TO 100 NS
- c) Rise time:
≤ 20 NS
- d) Fall time:
≤ 20 NS
- e) PRF: *MAX PW: 0 TO 50MHz*
PW ≤ 10NS: 0 TO 1KHz
- f) Jitter, stability:
OK
- g) Prime power:

① *POUT R_L = 50Ω*
20 US/DIV
100 VOLTS/DIV



120/240V
50-60Hz

② *POUT R_L = 50Ω*
200 NS/DIV
100 VOLTS/DIV

[Signature]