


PULSE GENERATOR
PERFORMANCE CHECK


Model: *PMMP-2-C-P*

S.N.: *5713*

Date: *DEC 3 1990*



*(A) 40 db ATTEN
- 10 VOLTS/DIV
PRF = 1.0 MHz*



*(B) AS (A) BUT 1.0 NS/DIV
(RISE TIME).*

- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 130 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*120 / 245V
50-60 Hz*

[Signature]