


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMR-3-PW-PS-T*
S.N.: *5719*
Date: *NOV 28 1990*



R_{in} = 50 Ω
PRF = 50 Hz

- a) Output signal amplitude: *0 TO +200 VOLTS*
- b) Pulse width: *TO 50 ns*
- c) Rise time: *0.1 TO 100 ns*
≤ 10 ns
- d) Fall time: *≤ 10 ns*
- e) PRF: *MAX PW: 0 TO 50 Hz*
PW ≤ 10 ns: 0 TO 1 kHz
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 Hz