


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2A-C-BSC1*

S.N.: *5743*

Date: *DEC 4 1990*



60 db ATTEN
100 VOLTS/DIV
PRF = 10 KHZ

- a) Output signal amplitude:
0 TO +170 VOLTS
- b) Pulse width: *TO 50 n*
- c) Rise time:
5 TO 100 NS
- d) Fall time:
≤ 2 NSEC
- e) PRF: *0 TO 20 KHZ*
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 HZ
- h) DELAY: *0 TO 7, 10 USEC*