


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ARR-3-PW-PS-P-TEL*

S.N.: *5744*

Date: *DEC 3 1990*



10. VOLTS/DIV
1.0 MS/DIV
PRF = 10 Hz
R_c = 3 Ω

- a) Output signal amplitude:
0 TO +21 VOLTS
- b) Pulse width: *TO 3 Ω*
1 TO 10 MS
- c) Rise time:
≤ 30 NS
- d) Fall time:
≤ 30 NS
- e) PRF: *MAX PW: 0 TO 10 Hz*
MIN PW: 0 TO 100 Hz
- f) Jitter, stability: *(10%) MAX*
DUTY CYCL.
OK
- g) Prime power:
120/240, 50
-60 Hz