


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *M4-M1-PN-SSB*

S.N.: *5746*

Date: *DEC 21 1990*



*Test 60 db ATTEN*



*Next 60 db ATTEN*

- a) Output signal amplitude:  
*± 100 VOLTS*
- b) Pulse width:  
*0.6 TO 2.0 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 ± 1.0 VOLTS*  
*200 mA max*

*[Signature]*