


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV6-3-C-N-0A-F01CB*

S.N.: *5752*

Date: *JAN 31 1991*



- a) Output signal amplitude:  
*-50, -75, -100, -150, -200, -250, -300, -350, -400, -500V;*
- b) Pulse width:  
*≤ 5 ns (1/2 power)*

c) Rise time:  
*NA*

d) Fall time:  
*NA*

e) PRF:  
*0 TO 20 KHz*

f) Jitter, stability:  
*OK*

g) Prime power:  
*120/240V*  
*50-60Hz*

*60 db ATTEN*

*≈ 100 VOLTS/DIV*

*PRF = 10 KHz*