

PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-A1-C-7-EBM2*

S.N.: *5767*


Date: *MAR 7 1991*

*SEE FOLLOWING
PAGE FOR
REQUIRED
WAVEFORMS*


- a) Output signal amplitude:
0 TO 6 AMPS
- b) Pulse width:
0.4 μ S (FIXED)
- c) Rise time:
20 TO 50 NS
- d) Fall time:
20 TO 50 NS
- e) PRF:
1 KHz (FIXED)
- f) Jitter, stability:
OK
- g) Prime power:
*120/200V
50-60 Hz*

[Signature]

AVR-A1-C-P-EBAR (SN 5767)




(A)



50 VOLTS/DIV
100 NS/DIV

(3 AMPS)

TR CONTROL
MID RANGE



(B)

AS (A) BUT
AMPLITUDE
REDUCED

2.0 VOLTS/DIV
100 NS/DIV

(100 mA)

TR CONTROL STILL
MID RANGE.

MAR 1 90
[Signature]