


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVR-E4-PS-P*

S.N.: *5842*

Date: *MAY 5 1991*



*60 dB ATTEN*

*100 VOLTS/DIV*

*PRF = 100 KHz*

- a) Output signal amplitude:  
*0 TO 700 VOLTS*
- b) Pulse width:  
*1 TO 5 NS*
- c) Rise time:  
*0.4 NS*
- d) Fall time:  
*0.4 NS*
- e) PRF:  
*0 TO 100 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 740 V*  
*50 - 60 Hz*