


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APB-9B-D-MU-MI*

S.N.: *5853*


Date: *MAY 30 91*



- a) Output signal amplitude:
0 TO + 200 mA
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 200 PS
- d) Fall time:
≤ 200 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

Ⓐ *APB-9B OUTPUT TO 50 Ω 40dB ATTEN ≈ 14 VOLTS*

*+15 VOLTS
300 mA (on PA)*



Ⓑ *MI OUTPUT OF PA-S1 MODULE ≈ 200 mA/DIV.*