


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVL-USA-CA-TP*

S.N.: *5873*

Date: *JUNE 26 1991*



- a) Output signal amplitude:  
*0 TO 100 VOLTS*
- b) Pulse width:  
*10 NS*
- c) Rise time:  
*≤ 1 NS*
- d) Fall time:  
*≤ 2 NS*
- e) PRF:  
*0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*15 V*  
*250 mA (max)*

*JH*