


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVK-ANP*

S.N.: *5880*

Date: *MAY 22 1991*



Ⓐ *40 db ATTEN*
10 VOLTS/DIV



Ⓑ *AS Ⓐ BUT 200 PS/DIV*
(RISE TIME)

- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
1 - 50 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
250 mA (MAX)

[Signature]