


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVG-1-PN-55A*

S.N.: *5889*

Date: *MAY 29 '91*



*Prot. 60 db atten
100 VOLTS/DIV*



*Prot. 60 db atten
100 VOLTS/DIV*

- a) Output signal amplitude:
*P: +250V
N: -250V*
- b) Pulse width:
≤ 5 nS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: *0 TO 50 KHz*
- f) Jitter, stability:
OK
- g) Prime power:

*+15 VOLTS
300 mA (max)*

[Signature]