


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVP-12-15-C-N*

S.N.: *5890*

Date: *MAY 31 1991*



Ⓐ *40dB ATTEN*  
*5 VOLTS/DIV*  
*PRF = 1.0 MHz*



Ⓑ *AS Ⓐ BUT 100 PS/DIV*  
*(RISE TIME)*

- a) Output signal amplitude:  
*0 TO -5V OCB*
  - b) Pulse width:  
*0.2 TO 4.0 NS*
  - c) Rise time:  
*≤ 60 PS*
  - d) Fall time:  
*≤ 135 PS*
  - e) PRF:  
*0 TO 1 MHz*
  - f) Jitter, stability:  
*OK*
  - g) Prime power:  
*120/240 V*  
*50-68 Hz*
- [Signature]*