


PULSE GENERATOR

PERFORMANCE CHECK

Model: *APR-3-PW-C-P-M1A3*


S.N.: *5901*

Date: *MAY 30 1991*



- a) Output signal amplitude: *0 TO +20 VOLTS*
- b) Pulse width: *TO 10 ns*
- c) Rise time: *2 TO 25 ns*
- d) Fall time: *≤ 50 ns*

ⓐ *5.0 ns/DIV, 10 VOLTS/DIV*
R_L = 10 Ω
PRF = 1.0 KHz



- e) PRF: *10 Hz TO 10 KHz*
(2.5? MHz)
- f) Jitter, stability: *DUTY*
CYCLE
OK
- g) Prime power: *120/240 V*
50-60 Hz

ⓑ *200 ns/DIV*
10 VOLTS/DIV
R_L = 10 Ω

[Signature]