


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMR-7B-C-AT-PN*


S.N.: *5904*

Date: *AUG 1 1991*



- a) Output signal amplitude: *0 TO ± 700 VOLTS*
- b) Pulse width: *0.1 TO 100 μS*
- c) Rise time: *≤ 50 NS*
0.5% MAX DUTY CYCLE
- d) Fall time: *≤ 50 NS*
- e) PRF: *0 TO 10 KHz,*
0.5% MAX DUTY
- f) Jitter, stability: *OK*
CYCLE
- g) Prime power: *120/240V*
50-60 Hz.

(A) Rout R_c = 50Ω
PRF = 50 Hz
≈ 200⁺ VOLTS/DIV



(B) Rout R_c = 50Ω
PRF = 10 KHz.
≈ 200⁺ VOLTS/DIV.

[Signature]