


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PH-121-PN-55C*

S.N.: *5917*

Date: *JUN 12 1991*



a) Output signal amplitude:  
*± 100 VOLTS.*

b) Pulse width:  
*0.6 TO 2.0 NS*

c) Rise time:  
*NA*


d) Fall time:  
*NA*

e) PRF:  
*0 TO 100 KHZ*

f) Jitter, stability:  
*OK*

g) Prime power:  
*+10.2 TO 14.0 VOLTS  
250 mA MAX.*

Ⓐ *Pass 60 db ATTEN  
50 KHZ PRF*



Ⓑ *Not 60 db ATTEN*

*AA*