


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV4-H1-PN-55C*

S.N.: *5918*

Date: *JUL 12 1991*



a) Output signal amplitude:  
*± 100 VOLTS*

b) Pulse width:  
*0.6 TO 2.0 NS*

c) Rise time:  
*NA*

d) Fall time:  
*NA*

e) PRF:  
*0 TO 100 KHz*

f) Jitter, stability:  
*OK*

g) Prime power:  
*+ 10.2 TO 14.0 VOLTS*  
*250 mA (max)*

*A* *P<sub>out</sub> 60 db ATTEN*  
*50 KHz PRF*



*B* *N<sub>out</sub> 60 db ATTEN*