


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV6-1-P

S.N.: 5925

Date: JUNE 4 1991



60 db ATTEN

100 VOLTS/DIV

PRF = 20 KHZ

- a) Output signal amplitude:  
0 TO +250 VOLTS
- b) Pulse width:  
≤ 4 NS
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 50 KHZ
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 VOLTS )  
250 mA (MAX)