


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PAK*
S.N.: *5935*
Date: *July 19 1991*



- a) Output signal amplitude:
0 TO ±100 VOLTS
- b) Pulse width: *(5-20)*
50 NS TO 1.00S
- c) Rise time:
< 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

R_L = 50 Ω
PRF = 20 KHz

120/240 V
50-60 Hz

[Signature]