


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-62-C-N-CAD-EWD*

S.N.: *5938*

Date: *JUNE 27 1991*



a) Output signal amplitude:

0 TO 50 VOLTS/DIV

b) Pulse width: *TO 50 n*

20 TO 200 NS

c) Rise time:

± 0.5 NS

d) Fall time:

± 0.5 NS

e) PRF:

0 TO 50 KHz

f) Jitter, stability:

OK


g) Prime power:

120/240 V


50-60 Hz

[Signature]

Ⓐ *50 dB ATTEN = 32 VOLTS/DIV*
PRF = 50 KHz



Ⓑ *AS Ⓐ BUT 1NS/DIV*
LEADING EDGE



Ⓒ *AS Ⓐ BUT 1NS/DIV*
FALLING EDGE