


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-107B-C-P-M*

S.N.: *5983*

Date: *SEPT 9 1991*



10 V/DIV

R_L = 6.5 Ω

100 NS/DIV

PRF = 20 kHz

- a) Output signal amplitude:
0 TO +2 AMP
- b) Pulse width: *(0 TO +60 us)*
- c) Rise time:
20 TO 200 NS
 ≤ 10 NS
- d) Fall time:
 ≤ 10 NS
- e) PRF:
0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50 to 1 Hz