

PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-2-AOA1

S.N.: 5988

Date: AUG 8 91



- a) Output signal amplitude:
-200 VOLT
- b) Pulse width: (TO 50ns)
10 TO 50NS
- c) Rise time:
2.5 NS
- d) Fall time:
5 NS
- e) PRF:
0 TO 2 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
+24 VOLTS DC
550 mA (MAX)

$R_L = 50 \Omega$
PRF = 1.0 KHZ

SEE FILE 4, PAGE 110,
OF CMS NO 8 FOR
GUIDANCE IN CONVERTING
TO HIGH IMPEDANCE
LOAD.