


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV6-1-P-4T1

S.N.: 5989

Date: JULY 30 1991



60 db ATTEN
.100 VOLTS/DIV
PRF ≈ 20 KHz.

- a) Output signal amplitude:
0 TO +250 VOLTS
- b) Pulse width: TO 50ns
- c) Rise time:
≤ 4 NS
- d) Fall time:
NA
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
250 mA (MAX)

h) TEMP RANGE:
THIS UNIT WAS
CHECKED AT
0°C, +24°C AND
+50°C.

[Signature]