


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AV-134-PS-R4-LIB1

S.N.: 6743

Date: FEB 3 1994



- a) Output signal amplitude:  
+ 1 VOLT (50 Ω)
- b) Pulse width:  
3 TO 6 NS
- c) Rise time:  
≤ 150 ps
- d) Fall time:  
≤ 300 ps
- e) PRF:  
0 TO 50 KHz
- f) Jitter, stability:  
OK

- g) Prime power:  
120 / 240 VOLTS  
50 - 60 Hz

- h) Propagation delay  
5.8 ± 0.2 NS

20 dB ATTEN  
- 1 VOLT/DIV