


PULSE GENERATOR
PERFORMANCE CHECK

Model: AN-151B-C

S.N.: 2030

Date: Jul 11 1997



- a) Output signal amplitude:
0 TO ± 200 VOLT
- b) Pulse width: ($R_L > 50K$)
SINE, SQUARE, TRIANGLE
+ DC OFFSET
- c) Rise time:
 $\leq 1 \mu S$
- d) Fall time:
 $\leq 1 \mu S$
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power: 120/240 V
50-60 Hz.

10 μS / DIV
50 V / DIV
 $R_L = 50K$.