


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-2-C-BE03-R5-P-0PI-M

S.N.: 8886

Date: Dec 20 1999



- a) Output signal amplitude: 0 TO +2 AMP ✓
- b) Pulse width: 17 NS ✓
- c) Rise time: ≤ 2 NS ✓
- d) Fall time: ≤ 2 NS ✓
- e) PRF: 0 TO 20 kHz ✓
- f) Jitter, stability: OK ✓
- g) Prime power: 120/240V, 50-60 Hz ✓

MONITOR OUT TO 40 dB
100V/DIV, 2 AMP/DIV
5 NS/DIV

AVO-2T-BE02B3 CONNECTED
VIA A 8 METER CABLE
IN459A IN SERIES WITH
1 Ω USED AS TEST LOAD.
PRF = 20 kHz

A handwritten signature or set of initials, possibly 'JTB', written in dark ink.