


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVR-E2-P-B*

S.N.: *9003*

Date: *OCT 14 1999*



- a) Output signal amplitude: *0 to +50V (to 50mV)*
- b) Pulse width: *20 to 200 ns*
- c) Rise time: *≤ 0.5 ns*
- d) Fall time: *≤ 0.5 ns*
- e) PRF: *0 to 50 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*

*50 dB ATTEN*  
*∴ ≈ 32 VOLTS/DIV*  
*TOP: 20 NS/DIV*  
*MID: 2 NS/DIV (RISE TIME)*  
*BOT: 2 NS/DIV (FALL TIME)*

*50-60 Hz*  
*[Signature]*