


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVMMH-2-C-M-PN*

S.N.: *9019*

Date: *NOV 18 1999*



- a) Output signal amplitude: *0 TO ± 30 V (R_L = 50Ω)*
- b) Pulse width: *≤ 400ps*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz

40 db ATTN
∴ .10 V/DIV/DIV
TOP PRF = 3.0 MHz
500 ps/DIV

BOT PRF = 25 MHz
5 ns/DIV