

PULSE GENERATOR
PERFORMANCE CHECK

Model: AV-110C-PS
S.N.: 9023
Date: NOV 2 1999

- a) Output signal amplitude: 0 to ± 100 V ($R_L \geq 10k$)
- b) GAIN: $\times 1$ to $\times 50$
- c) Rise time: $\leq 2 \mu s$
- d) Fall time: $\leq 2 \mu s$
- e) PRF: DC to 50 kHz
- f) Jitter, stability: OK
- g) Prime power: 120/240V
50-60Hz.



SN 9023.

V_{out}

V_{in}

