

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV02-A3-PS-T-UTB*

S.N.: *9024*



Date: *NOV 15 1999*

- a) Output signal amplitude: *0 to 60V to RL > 3. (200 amps max)*
- b) Pulse width: *100 ns to 2.0 μs*
- c) Rise time: *≤ 100 ns*
- d) Fall time: *≤ 100 ns*
- e) PRF: *0-25 Hz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V
50-60 Hz*

9024
 $R_c = 0.3 \Omega$

V_{out}

V_{in}



9024
 $R_c = 0.3 \Omega$

V_{out}

V_{in}

