


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANMM-2-PS TR-N*

S.N.: *9026*

Date: *OCT 4 1999*



- a) Output signal amplitude: *0 TO -5V (50nV)*
- b) Pulse width: *1 TO 10 NS*
- c) Rise time: *300ps TO 2.0 NS*
- d) Fall time: *300ps TO 2.0 NS*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *100 ± 10 uW*  
*50-60 dB*

*40 dB ATTEN*  
*-5 VOLTS/DIV*  
*TR SWITCH IN*  
*5 POSITIONS.*