


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *NHF-1*

S.N.: *9029*

Date: *NOV 8 1999*



a) Output signal amplitude:

*$\geq -10V$*

b) Pulse width:

*$\leq 130 ps$*

c) Rise time:

*NA*

d) Fall time:

*NA*

e) PRF:

*$100 \pm 5 MHz$*

f) Jitter, stability:

*OK*

g) Prime power:

*0.5 WATT (15V<sub>p-p</sub>)  
100 MHz CW.*

*40dB ATTN*

*10 V/DIV*

*TR 2 NS/DIV*

*BOT 200 ps/DIV.*