


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVHF-1
S.N.: 9031
Date: NOV 8 1999



- a) Output signal amplitude:
≈ -10 V₀₋₅
- b) Pulse width:
≤ 130 ps
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: 100 ± 5 MHz
- f) Jitter, stability: OK
- g) Prime power:
0.5 WATT (15 V_{p-p})
100 MHz CW

40 dB ATTEN
10 V₀₋₅ / DIV
TR 2 ns / DIV
BOT 200 ps / DIV

[Signature]