

PULSE GENERATOR

PERFORMANCE CHECK

Model: AVR-1-PW-PS-F

S.N.: 9037

Date: NOV 17 1999

- a) Output signal amplitude:
0 to +200V ($R_L \geq 50\Omega$)
- b) Pulse width:
50ns to 5.00µs
- c) Rise time:
 ≤ 10 ns
- d) Fall time:
 ≤ 10 ns
- e) PRF: 0 to 1 kHz.
- f) Jitter, stability:
OK
- g) Prime power: 100 ± 10V
50 ~~160~~ kHz

9037
 $R_L = 50 \Omega$

V_{out}

