


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANMP-2-P*

S.N.: *9039*

Date: *SEPT 29 1999*



- a) Output signal amplitude:  
*0 to +10 Volts*
- b) Pulse width: *(50ns)*  
*5 to 100ns*
- c) Rise time:  
*≤ 100ps*
- d) Fall time:  
*≤ 135ps*
- e) PRF:  
*0 to 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+ 24 Volts*  
*300mA max*

*40 dB ATTN*  
*10 Volts/div*  
*TOP 10 ns/div*  
*BOT 200ps/div (Rise Time)*

*[Signature]*