


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-MP-N*

S.N.: *9046*

Date: *SEPT 13 1999*



- a) Output signal amplitude:  
*- 40V TO 500MA*
- b) Pulse width:  
*2 TO 100 NS*
- c) Rise time:  
*≤ 350PS*
- d) Fall time:  
*≤ 2 NS*
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:

*+15 VDC  
200mA MA*

*50 dB ATTEN  
≈ 32 VOLTS/DIV  
PRF ≈ 300 KHz  
TOP DL ≈ 24"  
BOT DL ≈ 0"*